SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

D BECK COMPANY
200 Stonewall Blvd, Suite 5
Wrentham, MA 02093
Ian Beck Phone: 774 847 9869

CALIBRATION

Valid To: November 30, 2028 Certificate Number: 5942.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations[1, 5]:

I. Dimensional

Parameter/Equipment	Range	CMC[2, 4, 6] (±)	Comments
Optical Comparators[3] –			
Length	Up to 12 in	$(27 + 10L + 0.6R)$ µin	Reticle
Magnification	10X, 20X, 30X	72 µin	Mag checker
Squareness	Up to 4 in	50 µin	Angle block
Video Measurement Systems[3] – Linearity			
XY Axis	Up to 8 in Up to 18 in	76 µin 140 µin	Grids Linescale
Z Axis	Up to 6 in	140 µin	Gage blocks

[1] This laboratory offers commercial calibration service and field calibration service.

[2] Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

[3] Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

[4] In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches.

[5] This scope meets A2LA's *P112 Flexible Scope Policy.*

[6] In the statement of CMC, R is the resolution of the unit under test.

Accredited Laboratory

A2LA has accredited

D BECK COMPANY

Wrentham, MA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of A2LA's R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).

Presented this 6th day of October 2026.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 5942.01
Valid to November 30, 2028

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.